



PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

Portable Automatic Wastewater Samplers
Aquacell P2-COMPACT, Aquacell P2-COOLBOX,
Aquacell P2-MULTIFORM

Stationary Automatic Wastewater Samplers
Aquacell S50, Aquacell S100 & Aquacell S200 (Formally S2R),
Aquacell S310, S310H, S320, S320H & S320H+S50
Aquacell S510, S510H, S520 & S520H

Manufactured by:

Aquamatic Ltd
Irlam Business Centre
Soapstone Way, Irlam
Manchester
M44 6GP, UK

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards and Test Procedures for Continuous Water
Monitoring Equipment – Part 1, Version 4, April 2017 & EN 16479:2014**

Certification Range :

Lift height 0 to 7 metres (AC mains powered)
 0 to 5 metres (Battery powered)
Sample line diameter nominally 12.5mmID

Project No.: 80047861
Certificate No: Sira MC050059/19
Initial Certification: 24 June 2011
This Certificate issued: 08 July 2020
Renewal Date: 28 June 2025

Emily Alexander
Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*